














No.	Equipment to discard	Image	Asset tag	Current status	Decision: Retain (new location) or discard
1	Branson digital signifier		3000000442	PM to be done this month	Discard Comment: equipment not working; power supply module is faulty
2	Armfield FT40 mixer		3000000634	Not in use	Discard Comment: equipment not working
3	WAB CH-4132 Dynamill (BeadMill) <ul style="list-style-type: none"> • Include a low temperature circulator 		3000000780, 3000000781	Not in use	Discard

No.	Equipment to discard	Image	Asset tag	Current status	Decision: Retain (new location) or discard
4	SEW-Eurodrive BFT 50L Kettle		300000663	Not in use	Discard
5	Liquid nitrogen tank		No asset tag	Not in use	Discard
6	Lenze (AZ Tech) ESV751N01SMC Quadro Comil G97		300000633	Not in use	Discard

No.	Equipment to discard	Image	Asset tag	Current status	Decision: Retain (new location) or discard
7	Powder blender Alphe3		30001541	Not in use	Discard
8	accumet BASIC AB15 pH meter		3000000437	Not working	Discard
9	Sartorius PT-10 pH meter		No asset tag	Not working	Discard
10	VWR ph110 pH meter		No asset tag	Not working	Discard
11	Infrared/ type K Thermometer		No asset tag	Not working	Discard
12	Infrared/ type K Thermometer		No asset tag	Not working	Discard

No.	Equipment to discard	Image	Asset tag	Current status	Decision: Retain (new location) or discard
13	pH meter		No asset tag	Not working	Discard